

Indexed by

Scopus[®]**INTERNET OF THINGS BASED REAL-TIME VITAL SIGN MONITORING SYSTEM USING MOBILE APPLICATION**DOAJ
DIRECTORY OF
OPEN ACCESS
JOURNALSCrossref**Sugondo Hadiyoso***Telkom University, School of Applied Science, Bandung, Indonesia***Akhmad Alfaruq***Telkom University, School of Applied Science, Bandung, Indonesia***Rohmat Tulloh***Telkom University, School of Applied Science, Bandung, Indonesia*ROAD
DIRECTOR OF OPEN ACCESS
RESEARCHKoBSON**Yuyun Siti Rohmah***Telkom University, School of Applied Science, Bandung, Indonesia***Erwin Susanto***Telkom University, School of Electrical Engineering, Bandung, Indonesia*SCINDEKS
Srpski citatni indeksGoogle
Scholar**Key words:** *telehealth, IoT, vital sign, real-time, monitoring***doi:**10.5937/jaes0-28774**Cite article:**

Hadiyoso, S., Alfaruq, A., Tulloh, R., Rohmah, Y. S., Susanto, E. (2021) INTERNET OF THINGS BASED REAL-TIME VITAL SIGN MONITORING SYSTEM USING MOBILE APPLICATION, *Journal of Applied Engineering Science*, 19(3), 807-813, DOI:10.5937/jaes0-28774

Online access of full paper is available at: www.engineeringscience.rs/browse-issues

INTERNET OF THINGS BASED REAL-TIME VITAL SIGN MONITORING SYSTEM USING MOBILE APPLICATION

Sugondo Hadiyoso^{1,*}, Akhmad Alfaruq¹, Rohmat Tulloh¹, Yuyun Siti Rohmah¹, Erwin Susanto²

¹Telkom University, School of Applied Science, Bandung, Indonesia

²Telkom University, School of Electrical Engineering, Bandung, Indonesia

The development of telehealth technology in monitoring systems has been widely used to support applications in the health sector. The aim is to provide easy access for the people. One of the implications is a real-time monitoring system based on the Internet of Things (IoT) platform. Some health vital signs that focus on observation are electrocardiogram (ECG) signal, oxygen saturation, blood pressure, and heart rate, providing heart health information. In this study, an integrated system has been implemented, namely a vital sign distributed monitoring system through the internet network. The implemented system was able to acquire vital signs then send data to the internet cloud to be stored and processed further for real-time monitoring needs by interested parties. An Android-based application called iHealthVitalSign monitor can send, process, and represent data in numerical and graphical forms. The average delay for each packet delivery was 154.73 ms and complied with the ITU-T recommendations for real-time data transfer. Heart rate (HR) detection algorithms have been evaluated on real-time ECG signals. From the test results for more than 2100 beats, the average detection accuracy is 98.78%. With this proposed application, it is hoped that it can increase the penetration of telehealth services.

Key words: telehealth, IoT, vital sign, real-time, monitoring

INTRODUCTION

The development of telehealth technology is more sophisticated and practical has been widely used to support applications in the health sector [1]. One of them is the Wireless Sensor Network (WSN) technology for monitoring vital signs. Furthermore, WSN is linked to the widely accessible internet, and this technology is then known as the Internet of Things (IoT). The Internet of things is a dynamic network with intelligent interface capabilities that can be seamlessly integrated into global networks [2].

Some of the most important health vital signs that are of concern are ECG, oxygen saturation (SpO₂), blood pressure (BP), and heart rate (HR) that can provide information on heart health. Observation of these signals is sometimes carried out continuously, not limited to places and times, which means that it is not only done in health centers. Therefore, we need an integrated system that provides a solution for these challenges. One of them is making acquisition devices connected to the internet to facilitate data and information transfer.

From these reviews, it is clear that IoT technology will be widely used to facilitate access to health services, especially online monitoring applications with high mobility. Acceleration of research is needed to support the availability of these devices and systems in a relatively short time. Therefore, the proposed research that will be carried out is building a system (model and implementation) of healthcare monitoring distributed through a wide internet network.

Some researchers have built IoT-based vital sign monitoring systems, and they mostly focus on ECG moni-

toring. Research by Nguyen, et al. [3] has designed and implemented the healthcare monitoring system, which uses an IoT-based 6LoWPAN network. He researched to make a real-time monitoring system centralized with several network topologies for many users. Online applications for ECG monitoring based on PC or mobile devices were also reported in the study [4]–[8].

Other research by Junaid. et al. [9] has developed Remote Patient Monitoring. His study's focus is to develop an android application for monitoring ECG signals and data storage for uploading to the cloud to be accessed by medical parties. This study does not address multi-user applications. Implementation of a health monitoring system such as ECG, body temperature, or blood pressure with local or point to point access was also reported in the study [10]–[14]. Research by Sugondo. et al. have realized an android application to detect arrhythmias using Bluetooth as a media of communication between signal acquisition devices and smartphones [15]. However, the monitoring system still uses peer to peer topology to not be used for many users. Other studies have focused on developing mobile applications to detect cardiac abnormalities based on a single vital parameter, as reported in [16]–[18]. Therefore, this study proposes a real-time application to monitor various vital sign parameters. This application also supports multi-sensor and multi-user implementations.

This proposed research is developing the study: PC-based Wireless Vital Sign Monitors use ZigBee and Wireless 802.11 LANs as data transmitter [19], [20]. In this research, the signal acquisition device of ECG, PPG, and body temperature was realized following monitoring applications on PC point to point through ZigBee and WLAN. Several

*sugondo@telkomuniversity.ac.id

network topologies have also been made using ZigBee to transmit ECG signal data in real-time [21]. The mentioned research has not been widely accessible through the internet, so it still has limited access. Recently a web-based digital ECG signal monitoring application has been developed, but it has not been running in real time.

The confirmed research, as explained above, is the main background for the development of this proposed study. The advantages obtained from this study are the availability of an embedded system of health devices for the acquisition of vital human signals that are integrated with global internet networks. Through the website page or mobile application, signals or data can be represented through graphs and numbers so that further analysis can be carried out. In the future, this application can be a prototype or model that is applied to healthservice centers and personal use.

SYSTEM DESIGN AND IMPLEMENTATION

The implemented system is depicted in Fig. 1. The vital sign acquisition device is used to measure the ECG signal, blood pressure, oxygen saturation, and body temperature. The main processing unit will then process the data and arrange the order of measurement data to be forwarded to the access point via the Bluetooth module. This proposed system's access point is a mobile phone as a gateway between sensor devices and internet networks. By mobile phone, then the data is forwarded to the server. Users who have interests can access the data through the "iHealthVitalSign" application installed on mobile devices.




Figure 1: Implementation of the proposed vital sign monitoring system



Figure 2: Hardware of acquisition device




Figure 3: Hardware connection

Signal acquisition device

The signal acquisition device consists of a single lead ECG, blood pressure meter, oxygen saturation meter, and a thermometer, as shown in Fig. 2. The ECG Analog front end design implemented is based on previous research [15]. The oxygen saturation sensor uses a MAX30100 based sensor module inside. The blood pressure acquisition device uses a BP meter with serial data output. Sensor 18B20 is used to acquire body temperature. The microcontroller is applied as the main controlling unit for processing and transmitting the data. The diagram of hardware design in this study can be seen in Fig. 3.

Bluetooth interface

The Bluetooth module in this study is used to transfer data from the main control unit to the mobile phone, which acts as an access point and a client. We use the Bluetooth HC-06 module that works in slave mode and, on the other hand, Bluetooth on the mobile phone as a master. The activation and Bluetooth connection on the mobile phone is made automatically after the QR-Code authentication process is successful and is received by the server. Fig. 3 shows the wiring diagram between the microcontroller and the HC-06.

Client-server architecture

Online and real-time applications in this study are implemented using client-server-based network architecture. This architecture provides a connectivity model on networks that distinguishes computer functions as client and server. This control function is carried out by an operating system embedded in the cloud. The server is responsible for providing services to connected clients. This service is in the form of communication lines, user authentication, database, and data transfer. This architectural model's working principle is straightforward, where the server will wait for a request from the client, process, and provide output to the client. Meanwhile, the client will send a request to the server, wait for the process and receive the desired package. This study uses the Transmission Control Protocol/Internet Protocol (TCP/IP), while the operating system used is Ubuntu

Server 18.04. The hardware specifications used by the server are shown in Table 1.

Table 1: Specification of Server

Type	Specification
Processor	2-Core Processor
ProcessingMemory	2 GB Ram
Storage	40 GB SSD
TrafficBandwidth	2 TB

iHealth vital sign application

iHealthVitalSign (Fig. 4) is an Android application integrated with vital signal acquisition devices that display measurement results in graphical and numerical forms. Users can take measurements online and in real-time using this application via the internet network. The application was developed based on Android, with a size of 3.66 MB. The system is equipped with authentication based on QR-code to run the application as an initial activation. Where each device has a unique QR code that is registered on the database server. The iHealth application is also equipped with a user's medical record that can be used to view medical history. iHealth is supported by web-based applications to perform functions related to user settings (doctors and patients). An administrator can access new user input, delete users, select doctors, and manage medical records. Applications that support iHealth can be seen in Fig. 5.





Figure 4: iHealthVitalSign apps. (a) login page (b) display of a vital sign




Figure 5: iHealth Supporting Application

Heart rate detection

Heart rate (HR) is one of the vital parameters that can be implicitly calculated based on the ECG signal. The QRS complex on the ECG is an important component in heart rate detection. Real-time HR calculations using the Pan Tompkins method to provide good performance in QRS detection according to what was a report(ed in a study by Pan and Tompkins [22]. Pan Tomkins has adaptive properties for each ECG signal pattern, so this algorithm is very suitable to be applied without having to do previous learning [22]. All QRS detection processes are implemented on mobile phones. The stages of the QRS detection process on Pan Tompkins are as follows:

1. Band Pass Filter (BPF)

BPF is used to reduce the noise of artifacts, muscle, and baseline wanderers. We designed BPF with a frequency range of 5-15 Hz by LPF and HPF. The design of the LPF and HPF follows the following difference equations:

LPF (Fc = 15 Hz)

$$y(n) = 2y(n-1) - y(n-2) + x(n) - 2x(n-6) + x(n-12) \tag{1}$$

HPF (Fc = 5 Hz)

$$y(n) = y(n-1) - 1/32 x(n) + x(n-16) - x(n-17) + 1/32 x(n-32) \tag{2}$$

2. Derivation

Derivation is used to get the QRS slope information by implementing the following equation:

$$y(n) = 1/8 [2x(n) + x(n-1) - x(n-3) - 2x(n-4)] \tag{3}$$

3. Squaring Function

This step is used to makes all data points positive. The equation of this operation is

$$y(n) = x^2(n) \tag{4}$$

4. Moving-Window Integration

Window integration is applied to get the slope of the R wave. It is calculated from

$$y(n) = 1 / N [x(n - (N - 1)) + x(n - (N - 2)) + \dots + x(n)] \quad (5)$$

where N is the size of the sliding window

5. Thresholds

The threshold function works adaptively to determine the heartbeat. Pan Tompkins uses two thresholds (low and higher) to identify peaks of the signal.

RESULTS AND DISCUSSION

The real-time vital sign monitoring system developed is an integrated system between signal acquisition devices and software applications. All communication protocols are configured smoothly so that this system can work with good performance. We conducted several tests to determine the performance of the proposed system so that it can be applied to real applications. Tests carried out include data integrity, transfer delay, storage capacity, HR detection, and medical records.

Data integrity

Integrity is one of the parameters for conformity testing between the data sent by the device acquisition and the data displayed by the iHealth application. Although there is a guarantee by the TCP/IP protocol, there is still a possibility of errors. This test is done by observing and comparing the data sent by the device, the server's data, and the data displayed by the application. In this study, we made a protocol in terms of timing of sending data based on necessity in displaying the examination results. Samples of ECG waves are sent continuously with a frequency of 100Hz to keep the ECG graph can be displayed smoothly. Meanwhile, SpO2 and body temperature are sent with a time interval of 2.5 seconds and 45 seconds. The value of the blood pressure measurement results will be sent when the user presses the button on the device. Several examination results of vital signs taken from serial communication are shown in Table 2.

Table 2: Measurement results of BP, SpO2, and body temp. on 10 subject

Subject	Blood Pressure		Temp.	SpO2
	Sys.	Dias.		
01	138	87	35.25	97
02	143	90	35	97
03	144	90	35.13	96
04	114	75	35.2	98
05	99	67	35.19	98
06	118	85	35.32	97
07	117	86	35.25	94
08	125	93	35.3	97
09	101	67	35.22	97
10	96	67	35.31	97

no	id_user	systolic	diastolic	spo2	temp
1	15	138	87	97	35.25
2	15	143	90	97	35
3	15	144	90	96	35.13
4	11	114	75	98	35.2
5	11	99	67	98	35.19
6	15	118	85	97	35.32

Figure 6: Data on server




Figure 7: Displayed data on apps

From the results of observations in Table 2, Fig. 6 and 7, we conclude that the device's measurement results have consistent values both on the database server and iHealth application. The proposed system can send and represent the vital sign parameter on a server and iHealth applications during the testing. In conditions where the SpO2 data, body temperature, and blood pressure are simultaneously sent can cause errors. In this situation, the patient can measure blood pressure with activation on the device.

Transfer delay

The real-time monitoring system is one of the objectives of the proposed system. One of the parameters to guarantee the fulfillment of QoS in real-time data transfer applications is a delay. The developed systems are susceptible to delay because they are very dependent on existing cellular traffic. Based on ITU-T recommendations, the standard data requirements for real-time services are <250 ms. Delay testing is done using a cellular internet network with an upload/download service speed of 19.3/4.91 Mbps. Testing is also done using a fixed internet network based on fiber optic transmission with the upload/download service speed of 1.60/9.96 Mbps. Delay calculation based on the time interval between packages sent and received as in example Fig. 8. The results of delay testing for each scenario are shown in Fig. 9.

```

log_ihealth.txt -- Edited
2018-10-30 11:21:10.904 20621-20621/al.akh.apps.ihealth E/AkhAl: SpO2 Send: 2018-10-30 11:21:10.900
2018-10-30 11:21:11.014 20621-20621/al.akh.apps.ihealth E/AkhAl: SpO2 Read: 2018-10-30 11:21:11.010
2018-10-30 11:21:11.486 20621-20621/al.akh.apps.ihealth E/AkhAl: ECG Send: 2018-10-30 11:21:11.482
2018-10-30 11:21:11.746 20621-20621/al.akh.apps.ihealth E/AkhAl: ECG Read: 2018-10-30 11:21:11.743
2018-10-30 11:21:13.404 20621-20621/al.akh.apps.ihealth E/AkhAl: SpO2 Send: 2018-10-30 11:21:13.400
2018-10-30 11:21:13.513 20621-20621/al.akh.apps.ihealth E/AkhAl: SpO2 Read: 2018-10-30 11:21:13.509
2018-10-30 11:21:15.909 20621-20621/al.akh.apps.ihealth E/AkhAl: SpO2 Send: 2018-10-30 11:21:15.905
2018-10-30 11:21:16.062 20621-20621/al.akh.apps.ihealth E/AkhAl: SpO2 Read: 2018-10-30 11:21:16.058
2018-10-30 11:21:16.997 20621-20621/al.akh.apps.ihealth E/AkhAl: ECG Send: 2018-10-30 11:21:16.993
2018-10-30 11:21:17.221 20621-20621/al.akh.apps.ihealth E/AkhAl: ECG Read: 2018-10-30 11:21:17.217
    
```

Figure 8: Log of send and receive time




Figure 9: Packet Delay

Based on Fig. 9, as many as 30 data packet samples have been tested in the process of sending and reading data. The average delay of the test for each scenario is 154.73 ms. This means that it meets the standards of the recommended real-time data communication requirements from ITU-T. From these results, it can be concluded that the system built can be used in real-time monitoring applications.

Storage capacity

The database server is used to store patient medical examination data. In this research, we use storage media with a maximum capacity of 40 GB. We analyze to estimate the number of examinations that patients can perform. The data sample size for each transmission is shown in Table 3.

Table 3: Data Size each parameter

Param.	Number of Bytes	Ann.
Blood Pressure	12	depending on activation
SpO2+Temp	12	every 10 seconds
ECG	5	sample/second. with duration 5-10

From Table 3 above, then we assume that the total duration of the examination is 5 minutes. The total size of data sent, including blood pressure data with 2 times measurements, is:

Table 4: Total size of data for 5 minutes

Param.	Number of Bytes	Total (Bytes)
Blood Pressure	12	24
SpO2+Temp	12	360
ECG	5	1500
	Total	1884

From Table 4, we know the total size of data at each examination is close to 1884 Bytes. If we have 40 GB of storage, this system can store the examination data as much as $\approx 21 \times 10^6$.

Evaluation of HR detection

System performance testing in detecting heart rates is an important parameter because this is a valuable feature in applications. We evaluate the Pan-Tompkins algorithm for real-time HR detection through ECG signals. Table 5 shows the results of the experimental evaluation of HR detection.

We have evaluated more than 2100 beats from 30 ECG signals. Table 5 shows the performance of the algorithm that is applied to the mobile application. The average error is 1.22% due to false positives and false negative detection. The high magnitude of T wave or QRS peak, which tends to be negative in some people, can cause error detection.

Table 5: Results of HR detection

No.	Total HR	HR detected	%error
1	60	59	1,67
2	60	60	0
3	60	60	0
4	60	61	1,67
5	120	121	0,83
6	120	121	0,83
7	120	117	2,5
8	84	82	2,38
9	87	87	0
10	76	76	0
11	77	77	0
12	85	86	1,18
13	83	83	0
14	80	81	1,25
15	67	65	2,99
16	68	65	4,41
17	67	67	0
18	72	71	1,39
19	72	70	2,78
20	79	80	1,27
21	94	94	0
22	96	95	1,04
23	97	97	0
24	30	29	3,33
25	30	30	0
26	30	29	3,33
27	40	40	0
28	40	40	0
29	40	39	2,5
30	82	81	1,22
		Mean Error	1,22




Figure 10: Medical Record (a) Patient data (b) Sample of ECG recording

Medical record

A medical recording is a mandatory feature for health service applications and the telehealth monitoring system application. The system that we have developed is equipped with medical record applications on database servers managed through web pages and can be accessed through iHealth by a doctor's account. By default, when a user activates a vital sign device and is connected to a network, all measurement data is automatically stored in the server database until the device is deactivated. Each recording is labeled as a user ID, vital sign parameter, and time of data recording. The results of recording the patient's medical history can be seen through iHealth, as shown in Fig. 10. The doctor who has the authorization can access the patient's medical record, as shown in Fig. 11.

CONCLUSIONS

In this research, a real-time application for a vital sign monitoring system has been developed. The application was then called iHealth VitalSign, which is run on Android mobile. This system was able to measure blood pressure, ECG, SpO₂, temperature, and heart rate. iHealth was used to send data to the cloud server and represent data in numerical and graphical forms. This system uses the TCP/IP protocol to guarantee data sequences and perform error control functions.

The proposed system has been tested for its performance. Testing includes data integrity, transfer delay, storage capacity, HR detection, and medical records. The proposed system was able to store and represent medical check-up results on server applications and iHealth, which is by the results of device acquisition. The average delay for each packet delivery was 154.73 ms and met the ITU-T recommendations. This study also evaluated the HR detection algorithm through ECG signals. From tests that have been carried out on more than 2100 beats, it is known that the application generates a detection accuracy of 98.78%. The high amplitude of T





Figure 11: List of medical records on iHealth (a) record files (b) repeat ECG signals' playback

wave or QRS peak, which tends to be negative in some people, can cause error detection. The system that we have developed has been equipped with medical record applications on the database server and is fully accessible through iHealth by a doctor's role. The proposed system is expected to be applied for telehealth in health centers or personal use.

CONFLICTS OF INTEREST

All authors declare that they have no competing interests. All authors acknowledge and agree to the publication of this manuscript.

ACKNOWLEDGMENT

We would like to thank the Ministry of Research, Technology and Higher Education of the Republic of Indonesia, who has give us financial support through Penelitian Strategis Nasional Institusi scheme. Also for Telkom University for giving us full technical support through the laboratory and the research center.

REFERENCES

1. Dinesen, B., et al.(2016). "Personalized telehealth in the future: A global research agenda," J. Med. Internet Res., vol. 18, no. 3, 1–17.
2. Sundmaeker, H., De Saint-exupéry, A. (2010). Vision and Challenges for Realising the Internet of Things, no. March.

3. Gia, T. N., Thanigaivelan, N. K., Rahmani, A.M., Westerlund, T., Liljeberg, P., Tenhunen, H. (2015). Customizing 6LOWPAN networks towards internet-of-things based ubiquitous healthcare systems. NORCHIP 2014-32nd NORCHIP Conference: The Nordic Microelectronics Event. Institute of Electrical and Electronics Engineers Inc.
4. Komala, K., Kurian, M. Z., Shivannavar, Ashwini, S. (2009). Real time access & control of ECG signals using lab view based web browser. International Conference on Anti-counterfeiting, Security, and Identification in Communication.
5. Chai, J. (2013). The design of mobile ECG monitoring system. ICEIEC 2013 - Proceedings of 2013 IEEE 4th International Conference on Electronics Information and Emergency Communication, pp. 148–151.
6. Arun, C. S., Alexander, A. (2017). Mobile ECG monitoring device using wearable non contact armband. Proceedings of IEEE International Conference on Circuit, Power and Computing Technologies, IC-CPCT 2017, pp. 4–7.
7. Shi, H., Haoren Wang, Huang, Y., Zhang, Y., Liu, C. (2017). A Mobile Intelligent ECG Monitoring System Based on IOS. 2017 International Conference on Sensing, Diagnostics, Prognostics, and Control, pp. 149–153.
8. Ai, Z., Zheng, L., Qi, H., Cui, W. (2018). Low-Power Wireless Wearable ECG Monitoring System Based on BMD101. Proceedings of the 2018 37th Chinese Control Conference (CCC), Wuhan, China, 25–27 July 2018, pp. 7374–7379.
9. Mohammed, J., Lung, C.H, Ocneanu, A., Thakral, A., Jones, C., Adler, A. (2014). Internet of things: Remote patient monitoring using web services and cloud computing. In Proceedings - 2014 IEEE International Conference on Internet of Things, iThings 2014, 2014 IEEE International Conference on Green Computing and Communications, GreenCom 2014 and 2014 IEEE International Conference on Cyber-Physical-Social Computing, CPS 2014, pp. 256–263.
10. KIM, B. J., NOH, Y. H., JEONG, D. H. (2013) Implementation of Mobile based Multi-Function Health Monitoring System. In: Proceedings of the 2013 International Conference on IT Convergence and Security, Macao, December 2013. Piscataway, New Jersey: Institute of Electrical and Electronics Engineers, pp. 1-4.
11. Ungson, Y., Reyna, M. A., Bravo-Zanoguera, M. E. (2014). Desarrollo de un sistema de ECG ambulatorio basado en Arduino y telefonía móvil para transmisión inalámbrica. Pan American Health Care Exchanges, PAHCE.
12. Trivedi, S., Cheeran, A. N. (2017). Android based health parameter monitoring. Proceedings of the 2017 International Conference on Intelligent Computing and Control Systems, ICICCS 2017, pp. 1145–1149.
13. Wannenburg J., Malekian, R. (2015). Body Sensor Network for Mobile Health Monitoring, a Diagnosis and Anticipating System. IEEE Sens. J., vol. 15, no. 12, pp. 6839–6852.
14. Harmah, D. J., Kathirvelu, D. (2013). An ubiquitous miniaturized android based ECG monitoring system. 2013 International Conference on Emerging Trends in Computing, Communication and Nanotechnology (ICE-CCN), pp. 117–120.
15. Hadiyoso, S., Usman, K., Rizal, A. (2015). Arrhythmia Detection Based on ECG Signal Using Android Mobile for Athlete and Patient. International Conference on Information and Communication Technology (ICoICT) Arrhythmia, pp. 166–171.
16. Mena, L. J., et al. (2018). Mobile Personal Health Monitoring for Automated Classification of Electrocardiogram Signals in Elderly. Comput. Math. Methods Med., vol. 2018, pp. 1–9.
17. Sukanesh, R., Veluchamy, S., Karthikeyan, M. (2014). A Portable Wireless ECG Monitoring System using GSM Technique with Real Time Detection of Beat Abnormalities. Int. J. Eng. Res., vol. 3, no. 2, pp. 108–111.
18. Bazi, Y., Al Rahhal, M. M., Alhichri, H., Ammour, N., Alajlan, N., Zuair, M. (2020). Real-Time Mobile-Based Electrocardiogram System for Remote Monitoring of Patients with Cardiac Arrhythmias. Int. J. Pattern Recognit. Artif. Intell., vol. 34, no. 10, pp. 1–24.
19. Rizal, A., Suryani, V., Jondri, Hadiyoso, S. (2014). Development of Wireless Patient's Vital Sign Monitor Using Wireless LAN (IEEE.802.11.b/g) Protocol. nt. J. Electr. Comput. Eng., vol. 4, no. 6, pp. 893–901.
20. Fikar, A. R., Rahman, A., Rizal, A., Hadiyoso, S., Jondri. (2014). PC Based Wireless Vital Sign Monitor Using Zigbee Communication. The 15th International Conference on Biomedical Engineering, pp. 904–907.
21. Hadiyoso, S., Aulia, S. (2014). Multipoint to Point EKG Monitoring Berbasis ZigBee. Seminar Nasional Aplikasi Teknologi Informasi (SNATI), pp. 35–39.
22. Pan, J., Willis, J. (1985). A Real-Time QRS Detection Algorithm. IEEE Trans. Biomed. Eng., vol. BME-32, no. 3, pp. 230–236.

Paper submitted: 08.10.2020.

Paper accepted: 01.03.2021.

This is an open access article distributed under the CC BY 4.0 terms and conditions.