

EVALUATION OF AUTOMATED HEMATOLOGY ANALYZER DYMIND DH76 COMPARED TO SYSMEX XN 1000 SYSTEM

PROCENA AUTOMATSKOG HEMATOLOŠKOG ANALIZATORA DYMIND DH76 U POREĐENJU SA SYSMEX XN 1000 SISTEMOM

Milena Velizarova^{1,2}, Teodora Yacheva², Mariana Genova^{1,2}, Dobrin Svinarov^{1,2}

¹Department of Clinical Laboratory, Faculty of Medicine, Medical University of Sofia, Sofia, Bulgaria

²Clinical Laboratory & Clinical Pharmacology, UMBAL Alexandrovska, Sofia, Bulgaria

Summary

Background: DYMIND DH76 (DYMIND BIOTECH, China) is a new automated hematology system designed to provide CBC count, including a 5-part WBC differential count, and its analytical performance should be assessed before adoption for clinical use.

Methods: The analyzer was evaluated according to the International Council for Standardization in Haematology guideline. The purposes of this study were to assess its analytical performance in comparison to SYSMEX XN 1000 hematology analyzer currently used in our laboratory, as well as to compare the automated and manual WBC differential.

Results: Within-run precision in all concentration ranges was very good with coefficients of variation (CVs) between 0.02% and 2.5% except for platelets over $500 \times 10^9/L$ (CV 9.5%). Within-batch imprecision showed CVs lower the declared deviation ranges. Accuracy (defined as trueness) was excellent for all CBC and white cell differential parameters, compared with the state of the art%. Linearity was confirmed with excellent regression coefficients (0.999–1.000), even in the lowest values, and carryover was $\leq 1\%$. Comparison between DYMIND DH76 and SYSMEX XN 1000 was also very good with correlation coefficients (R^2) for WBC (1.000), RBC (0.999), hemoglobin (0.999) and PLT over $50 \times 10^9/L$ (0.994) and R^2 was lower but still acceptable (0.910) for PLT $< 50 \times 10^9/L$. R^2 for neutrophils, lymphocytes, eosinophils, basophils, and monocytes were 0.974, 0.982, 0.957, 0.625, and 0.836, respectively, in the comparison between the manual and DYMIND DH76 automated differential WBC counts.

Kratak sadržaj

Uvod: DYMIND DH76 (DYMIND BIOTECH, Kina) je novi automatski hematološki sistem napravljen da pokaže broj CBC, uključujući petodelni diferencijalni broj WBC. Pre ulaska u kliničku upotrebu, potrebno je da se ispituju njegove analitičke performanse.

Metode: U ocenjivanju analizatora korišćene su smernice Međunarodnog saveta za standardizaciju u hematologiji. Svrha ove studije je bila procena njegovih analitičkih performansi u odnosu na hematološki analizator SYSMEX XN 1000 koji se trenutno koristi u našoj laboratoriji, kao i da se uporedi automatski i ručni WBC diferencijal.

Rezultati: Preciznost u toku ispitivanja u svim rasponima koncentracija je bila vrlo dobra sa koeficijentima varijacije (CV) između 0,02% i 2,5%, osim za trombocite preko $500 \times 10^9/L$ (CV 9,5%). Nepreciznost unutar serije je pokazala da CV smanjuje deklarirani opseg odstupanja. Tačnost (definisana kao istinitost) je bila odlična za sve CBC i diferencijalne parametre belih ćelija, u poređenju sa procentom vrhunske tačnosti. Linearnost je potvrđena odličnim koeficijentima regresije (0,999–1,000), čak i u najnižim vrednostima, a prenos je bio $\leq 1\%$. Poređenje između DYMIND DH76 i SYSMEX XN 1000 je takođe bilo vrlo dobro sa koeficijentima korelacije (R^2) za WBC (1,000), RBC (0,999), hemoglobin (0,999) i PLT preko $50 \times 10^9/L$ (0,994), i R^2 je bio niži, ali i dalje prihvatljiv (0,910) za PLT $< 50 \times 10^9/L$. R^2 za neutrofile, limfocite, eozinofile, bazofile i monocite je bio redom 0,974; 0,982; 0,957; 0,625 i 0,836 u poređenju između ručnog i DYMIND DH76 automatskog diferencijalnog broja WBC.

Address for correspondence:

Milena G. Velizarova
1 G. Sofiiski str., Sofia, Bulgaria
Phone number: +35929230926,
e-mail: milena.velizarova@gmail.com

List of abbreviations: ICSH, International council for standardization in hematology; ALY, abnormal lymphocytes; LIC, large immature cells; P-LCR, platelet large cell ratio; P-LCC, platelet large cell count; STAT (from Latin statum), without delay, urgent; CV, correlation coefficient; SD, standard deviation; CBC, complete blood cell count.

Conclusions: With excellent analytical performance and acceptable comparative analysis, DYMIND DH76 hematology analyser covered the predefined international standards and requirements and is fully appropriate for clinical application.

Keywords: hematology analyser DYMIND DH76, performance evaluation

Introduction

Clinical laboratories need to vigorously assess the performance characteristics and reliability of each analytical instrument before adoption for clinical use. DYMIND DH76 (DYMIND BIOTECH, China) is a new automated hematology analyser designed to report 29 parameters, including a 5-part WBC differential count, with a capacity for the analysis of 80 samples (CBC/DIFF) per hour. At the same time, it provides an enumeration of abnormal lymphocytes – ALY (% , #), large immature cells – LIC (% , #), platelet large cell ratio (P-LCR), and platelet large cell count (P-LCC). The analyzer works with anticoagulated (K_2 EDTA or K_3 EDTA) whole blood (venous and capillary), aspiration volume is 20 μ L (mode CBC/DIFF), and pre-diluted samples could be processed in a special operation procedure. There is a possibility to interrupt the running series for urgent analyses (STAT function). DYMIND DH76 uses the impedance method with hydrodynamic focusing for counting WBC/BAS, RBC, and PLT, the non-cyanide colorimetric method for hemoglobin measurement, and semiconductor laser-based flow cytometry with light scattering as the principle for differential and total leukocyte count. Simultaneously, the analyzer provides information on the distribution of cells in three histograms, two 2D- and one 3D-scattergrams. Analytical characteristics declared by the manufacturer are presented in *Tables I – III*. DYMIND DH76 is able to generate flags in the presence of morphologically abnormal WBC, RBC, and PLT. The aim of this study was to assess the analytical performance of DYMIND DH76 with respect to: 1) manufacturer declared specifications (precision, reproducibility, carryover, and linearity); 2) comparison of DYMIND DH76 with the hematology analyzer SYSMEX XN 1000 currently used in our laboratory; 3) comparison between WBC differential count results obtained by DYMIND DH76 and microscopic morphologic («manual») examination. SYSMEX XN 1000 system (Sysmex, Kobe, Japan) determines erythrocyte and platelet counts by electrical impedance method and is able to perform an additional optical platelet measurement in the reticulocyte channel. For the WBC and differential counts, SYSMEX XN 1000 counter uses the flow cytometry method with fluorescent detection. WBC count is reported from the »WBC/BASO« channel, and control WBC data are generated from the independent »DIFF« channel. By comparing results produced by DYMIND DH76 with

Zaključak: Sa odličnim analitičkim performansama i prihvatljivom uporednom analizom, hematološki analizator DYMIND DH76 je pokrivaio unapred definisane međunarodne standarde i zahteve i u potpunosti je pogodan za kliničku primenu.

Ključne reči: hematološki analizator DIMIND DH76, procena performansi

results obtained from an automated system based on different analytical principles and with manual differential WBC count, the performance of the new instrument would be reliably proven.

Materials and Methods

Design of the study

In the present investigation, a total of 250 peripheral blood samples were analysed. For study purposes, fresh human whole blood samples, anticoagulated with K_2 EDTA, were used and processed no more than 2h after blood sampling. The blood samples were stored at room temperature to the time of analysis. Evidence of visible clots or samples with insufficient volume was a reason to probe rejection. Samples were selected to cover normal conditions and a very broad range of different types of the underlying pathology and thus encompassed reference ranges, low and high results.

In this study, only DYMIND-specified reagents, controls, and calibrators were used, according to the manufacturer's instructions. Analytical characteristics of DYMIND DH76 and the comparison with SYSMEX XN 1000 were evaluated according to International Council for Standardization in Haematology (ICSH) guideline (1).

Methods

Patient samples from the routine workflow, randomly selected in abnormally low, reference, and abnormally high analytical ranges were measured 10 consecutive times to assess reproducibility. For each hematologic parameter, the mean, SD, and CV% were calculated.

A single measurement repeated each day for a period of 30 days of the stabilized quality control material »CBC-5DMR Hematology Controls« – Normal (lotNo: BC1611), supplied by the manufacturer, was used to measure the total between-day precision for all included parameters. For each hematologic parameter mean value, SD and CV% were calculated.

Accuracy (assessed as trueness) was used to describe the closeness of a set of measurements to the true value (2, 3). The use of a »true value« for the

CBC is hard to apply in daily laboratory practice. For this reason, we studied the closeness of mean quality control material results, obtained by the analyses of quality control blood samples during the 30-day-period, to »target« values declared by the manufacturer of the quality control material. The percent deviation (d%) of the mean values from the »target« value was calculated for each parameter. Based on the current literature, our data were compared with state of the art for accuracy%.

Carryover was defined as the amount of analyte carried by the analyzer from one sample measurement into the subsequent measurement (1, 2, 4). It was mainly of importance for carryover from high to low concentrations of Hb, RBC, WBC, and platelets. The carried biological material for HGB, RBC, WBC, and PLT was evaluated. The percentage carryover was assessed by analyzing a sample with a high concentration three times (H1, H2, H3) followed by analyzing a sample with a low concentration three times (L1, L2, L3). Percentage carryover is calculated as follows: $\% \text{Carryover} = \frac{L1-L3}{H3-L3} \times 100$. For WBC, the low and the high values were $1.16 \times 10^9/\text{L}$ and $195.97 \times 10^9/\text{L}$, for RBC – $2.68 \times 10^{12}/\text{L}$ and $7.25 \times 10^{12}/\text{L}$, for HGB – 86 g/L and 208 g/L, and for PLT – $11 \times 10^9/\text{L}$ and $1787 \times 10^9/\text{L}$, respectively.

The evaluation of linearity showed the ability of the hematology analyzer to provide a result that was proportional to the analyser measured over a defined concentration range (1, 2, 4). For clinical purposes, the linear correlation between theoretical values and corresponded practical results, obtained by DYMIND DH76, was calculated especially for WBC, RBC, HGB, and PLT in low concentration ranges by the conduction of 1/2, 1/4, 1/8, 1/16 serial dilutions:

- The patient sample with the initial PLT value of $70 \times 10^9/\text{L}$ was diluted 1/2, 1/4, 1/8, 1/16 times to theoretical value of $4.3 \times 10^9/\text{L}$.
- The patient sample with the initial WBC value of $3.9 \times 10^9/\text{L}$ was diluted 1/2, 1/4, 1/8, 1/16 times to a theoretical value of $0.24 \times 10^9/\text{L}$.
- The patient sample with the initial RBC value of $4.66 \times 10^{12}/\text{L}$ was diluted 1/2, 1/4, 1/8, 1/16 times to a theoretical value of $0.29 \times 10^{12}/\text{L}$.
- The patient sample with the initial HGB value of 120 g/L was diluted 1/2, 1/4, 1/8, 1/16 times to a theoretical value of 7.5 g/L.

Dilutions were performed manually with the equipment solvent DIL-A Diluent (DYMIND BIOTECH, China). Dilutions were homogenized, then they were analyzed in duplicate. The test results were graphed and statistically analysed.

A comparison of the results from the evaluated DYMIND DH76 system with those obtained by the current hematology analyser SYSMEX XN 1000 was made for routine normal and abnormal samples in the laboratory.

Comparison between the manual and automated differential count of leukocytes was assessed. For the leukocyte differential count, the microscopic evaluation of a slide is currently regarded as the reference method (1, 5). The manual differential count was observed by two independent qualified researchers in 400 cells per May-Grünwald-Giemsa stained slides at an optical microscope.

All results obtained during the study, including those of quality control materials, were classified, recorded, and stored in the analyzer software program and in computer program Excel worksheets (Windows office). Printed results for all samples analyzed on both hematology analyzers were labeled and collected.

Statistical analysis

All statistical analyses were performed using IBM SPSS Statistics version 19.0 (IBM Corporation, Armonk, NY, USA). Patient sample correlations were calculated using Passing-Bablok regression and a difference comparison plot from the concordance study samples that were within the reportable range on both platforms. Pearson's correlation coefficient was employed to estimate linear relationships between the variables.

Results

Results from within-run imprecision studies of blood samples on the DYMIND DH76 hematology system are shown in *Table I*.

Results from between-batch precision studies, obtained by the analysis of quality control material in the normal range, are shown in *Table II*.

Accuracy was measured as trueness and was used to evaluate the closeness of a mean of measurements to »target« values of quality control material. The percent deviation (d%) data were compared with state of the art for accuracy% (*Table III*).

Linearity in low concentration ranges was verified for WBC, RBC, PLT, and HGB, and for all of them, it presented an excellent correlation coefficient (from 0.998 to 1.000) between expected theoretical and obtained values (*Table IV, Figure 1*).

Carryover was calculated according to the applied formula. For all of the analyzed parameters, carryover was less than 0.5% (*Table IV*).

Table I Within-run imprecision results.

Parameter	\bar{x}	SD	CV%	CV% Limits of acceptable imprecision*	CV% limits of declared imprecision**
WBC normal level ($3.5\text{--}10.5 \times 10^9/\text{L}$)	4.8	0.07	0.05	2.5	2.0
WBC low level ($<3.5 \times 10^9/\text{L}$)	2.9	0.04	0.04	6.0	5.0
WBC high level ($>11 \times 10^9/\text{L}$)	25.3	0.25	0.20	1.5	5.0
NEU (%)	59.6	0.51	0.39	-	4.0
LYM (%)	31.0	0.69	0.57	-	3.0
MON (%)	6.3	0.46	0.36	-	2.0
EOS (%)	2.8	0.22	0.18	-	1.5
BAS (%)	0.15	0.07	0.06	-	0.8
NEU ($\times 10^9/\text{L}$)	4.2	0.12	2.4	2.5	-
LYM ($\times 10^9/\text{L}$)	2.1	0.07	3.4	3.5	-
MON ($\times 10^9/\text{L}$)	0.6	0.05	7.8	8.5	-
EOS ($\times 10^9/\text{L}$)	0.2	0.07	2.8	10	-
BAS ($\times 10^9/\text{L}$)	4.8	0.13	2.9	20	-
RBC normal level ($3.50\text{--}6.00 \times 10^{12}/\text{L}$)	4.67	0.06	0.05	1.1	1.5
RBC low level ($<3.5 \times 10^{12}/\text{L}$)	1.61	0.04	0.03	-	-
RBC high level ($>6.00 \times 10^{12}/\text{L}$)	9.76	0.09	0.07	-	-
HGB normal level (120–180 g/L)	141.4	0.69	0.60	0.9	1.5
HGB low level (<100 g/L)	43.3	0.48	0.42	-	-
HGB high level (>180 g/L)	186.9	1.09	0.92	-	-
HCT (L/L)	0.43	0.01	0.01	1.2	-
MCV (fl)	93.0	0.58	0.33	0.6	1.0
MCH (pg)	30.2	0.38	0.25	1.1	-
RDW-CV (%)	13.3	0.07	0.05	2.0	-
PLT normal level ($130\text{--}400 \times 10^9/\text{L}$)	233.1	3.66	2.5	3.0	4.0
PLT low level ($<30 \times 10^9/\text{L}$)	2.7	1.05	0.9	4.5	8.0
PLT high level ($>500 \times 10^9/\text{L}$)	570.5	16.62	9.4	3.0	8.0

WBC, white blood cell count; RBC, red blood cell count; HGB, hemoglobin; HCT, hematocrit; MCH, mean corpuscular hemoglobin; MCV, mean corpuscular volume; RDW, red cell distribution width; PLT, platelets; NEU, neutrophils; LYM, lymphocytes; MON, monocytes; EOS, eosinophils; BAS, basophils; CV, coefficient of variation; SD, standard deviation;

*Limits of acceptable within-run imprecision, based on current literature (5, 16, 18, 19);

**CV% declared imprecision by the DYMIND DH76 system manufacturer.

Agreement between the obtained results from DYMIND DH76 and SYSMEX XN 1000 for complete cell count and differential WBC count was determined on 186 routine samples. Correlation statistics are presented in *Table V*. On the regression scatter plots, DYMIND DH76 results were dependent (y)

variables, and SYSMEX XN results were the independent (x) variables (*Figure 2*).

The correlation of the automated differential count of WBC with the manual method is shown in *Table VI*, and the difference plot estimation is shown in *Figure 3*.

Table II Between-batch imprecision results (quality control material in the normal range).

Parameter	\bar{x}	SD	CV%	CV% Limits of acceptable imprecision*	CV% limits of declared imprecision**
WBC ($\times 10^9/L$)	7.17	0.18	2.5	1.5	2
NEU (%)	58.8	1.32	2.2	2.5	-
LYM (%)	29.4	1.03	3.5	3.5	-
MON (%)	9.06	0.6	6.6	8.5	-
EOS (%)	2.72	1.03	1.9	10	-
BAS (%)	67.0	1.32	2.2	20	-
NEU ($\times 10^9/L$)	4.2	0.12	2.8	2.5	-
LYM ($\times 10^9/L$)	2.1	0.08	3.7	3.5	-
MON ($\times 10^9/L$)	0.6	0.05	7.8	8.5	-
EOS ($\times 10^9/L$)	0.19	0.07	1.5	10	-
BAS ($\times 10^9/L$)	4.8	0.14	2.9	20	-
RBC ($\times 10^{12}/L$)	4.99	0.06	1.3	1.1	1.5
HGB (g/L)	139.8	1.4	1.0	1.0	1.5
HCT (L/L)	0.45	0.005	1.0	1.4	-
MCV (fl)	90.8	1.09	1.2	0.8	-
MCH (pg)	27.9	0.46	1.7	1.5	-
RDW-CV (%)	16.9	0.17	1.0	2.0	-
PLT ($\times 10^9/L$)	243.9	14.9	6.1	3.0	4
MPV (fl)	9.3	0.16	1.7	2.5	-

WBC, white blood cell count; RBC, red blood cell count; HGB, hemoglobin; HCT, hematocrit; MCH, mean corpuscular hemoglobin; MCV, mean corpuscular volume; RDW, red cell distribution width; PLT, platelets; NEU, neutrophils; LYM, lymphocytes; MON, monocytes; EOS, eosinophils; BAS, basophils; MPV, mean platelet volume; CV, coefficient of variation; SD, standard deviation.

*Limits of acceptable within-run imprecision, based on current literature (5, 16, 18, 19).

**CV% declared imprecision by the DYMIND DH76 system manufacturer.

Table III Analytical accuracy (measured as trueness) on DYMIND DH 76 system, compared with state of the art for accuracy %.

	n	x_0	\bar{x}	d%	State of the art (%)*
WBC ($\times 10^9/L$)	30	6.88	7.17	-4.2	4.4
RBC ($\times 10^{12}/L$)	30	4.86	4.99	-2.7	3.2
HGB (g/L)	30	138	139.8	-1.3	1.3
HCT (L/L)	30	0.43	0.45	-4.7	1.8
PLT ($\times 10^9/L$)	30	247	243.9	1.3	6.4
MCV (fl)	30	90	90.8	-0.9	2.0
NEU ($\times 10^9/L$)	30	4.3	4.2	2.3	3.2
LYM ($\times 10^9/L$)	30	2.12	2.1	0.5	5.0
MON ($\times 10^9/L$)	30	0.70	0.6	14.3	15
EOS ($\times 10^9/L$)	30	0.17	0.2	-11.8	13
BAS ($\times 10^9/L$)	30	4.65	4.8	-3.2	32*

State of the art for accuracy is based on the current literature (8, 20).

Table IV Carryover (%) and linearity (in low concentration range), DYMIND DH 76.

	HGB (g/L)	RBC ($\times 10^{12}/L$)	WBC ($\times 10^9/L$)	PLT($\times 10^9/L$)
Carryover (%)	0%	0.4%	0.02%	0%
Declared carryover (%)	0.6%	0.5%	0.5%	1%
Linearity (r)	1.000	1.000	0.999	0.999

WBC, white blood cell count; RBC, red blood cell count; HGB, hemoglobin; PLT, platelets.




Figure 1 Linearity graph plots of HGB, RBC, WBC, and PLT.




Figure 2 Correlation of complete blood count and white cell differential count results from the DYMIND DH 76 system compared with results from the SYSMEX XN 1000 system.

Table V Correlation results from the concordance studies.

Parameter	r	Slope	Y-intercept	SYSMEX XN1000 Mean	DYMIND DH76 Mean	Bias
WBC	1.000	1.02	-0.18	16.3	16.5	-0.2
WBC $3.5 \times 10^9/L$, n=30	0.994	1.01	-0.04	2.14	2.11	0.03
RBC	0.999	0.97	0.09	4.05	4.02	0.03
RBC $3.5 \times 10^{12}/L$, n=60	0.996	0.98	0.07	3.1	3.09	0.01
HGB	0.999	1.02	-2.10	116.9	117.3	-0.4
HGB $100g/L$, n=60	0.996	1.04	-4.0	83.5	83.1	0.4
Hct	0.995	0.99	0.01	0.365	0.367	-0.002
MCV	0.987	0.99	2.20	92.4	94.1	-1.6
MCH	0.980	0.95	1.60	29.3	29.5	-0.2
MCHC	0.952	0.92	24.6	313.4	311.3	2.1
RDW-CV%	0.974	0.94	1.11	15.6	15.9	-0.3
PLT	0.994	0.97	3.1	234.7	230.7	4.0
PLT $50 \times 10^9/L$, n=30	0.910	1.25	0.8	19.3	23.2	-3.9
% Neutrophils	0.998	0.99	1.17	56.9	57.5	-0.6
% Lymphocytes	0.997	0.99	1.26	30.2	31.2	-1.0
% Eosinophils	0.992	0.99	0.14	2.0	2.1	-0.1
% Basophils	0.855	1.00	-0.16	0.4	0.2	0.2
% Monocytes	0.965	0.95	-1.15	7.8	6.3	1.5

WBC, white blood cell count; RBC, red blood cell count; HGB, hemoglobin; HCT, hematocrit; MCV, mean corpuscular volume; MCH, mean corpuscular hemoglobin; MCHC, mean corpuscular hemoglobin concentration; RDW, red cell distribution width; PLT, platelets.

Table VI Correlation of white blood cell differential results from the DYMIND DH76 compared with manual white blood cell differential counts.

Parameter	r	Slope	Y-intercept	Manual mean	DYMIND DH76 Mean	Bias
% Neutrophils	0.974	0.97	2.32	60.4	60.7	-0.3
% Lymphocytes	0.982	0.97	0.51	30.3	29.9	0.4
% Eosinophils	0.957	0.99	0.03	2.94	2.96	-0.02
% Basophils	0.625	0.44	0.13	0.20	0.17	0.03
% Monocytes	0.836	0.89	0.92	6.68	6.85	-0.17




Figure 3 Correlation analysis (A) and difference plot estimation (B) of white blood cell differentials from the DYMINDDH 76 system compared with manual differentials.

Discussion

In the current study, we evaluated the analytical performance characteristics (accuracy, precision, carryover, linearity in low concentration range) of the DYMINDDH 76 automated hematology system. At the same time, we performed a comparison between the studied DYMINDDH 76 and SYSMEX XN 1000 hematology analyzer, used routinely in our hospital laboratory.

The scope of examined hematological parameters with the purpose of evaluating precision depends largely on the type of laboratory which will use the hematology analyzer. For the needs of hospital laboratories, it is necessary to assess not only the reference area but also pathologically low and abnormally high areas of determination.

The determination of within-run imprecision was important to show the analyzer quality in test conduction and confirm the good repeatability of the analyzing samples with minimal appropriate differences. Within-run precision for the DYMINDDH76 system in all evaluated concentration ranges was very good and met the DYMINDDH76 specifications. Our results indicated low variation coefficients for within-run imprecision in abnormally low and abnormally high white blood cell count, hemoglobin concentration, hematocrit, and platelet count samples. In contrast to our expectations, platelet count of $2.7 \times 10^9/L$ showed within-run imprecision of 0.9%, which was better than the previously documented studies and the declared deviation range (<10%) (4, 6, 7).

Between-batch precision was performed with a quality control material supplied by the manufacturer.

We aimed to ensure the stability of the biological material during these 30 days of the study, and a quality control material was our appropriate choice. All coefficients of variation were excellent and were up to declared deviation ranges. The differential count of monocytes presented high CV% in the analyzed normal quality control level may be due to increased morphological variability of monocyte cells (9). Maciel et al. (10) documented at the same manner the precision problem of monocyte count in hematology analyzers; however, the clinical significance of this imprecision is low.

The percent deviation data (d%) results, obtained from the accuracy study, were compared with state of the art for accuracy% from current literature (8, 20). Accuracy (defined as trueness) of the DYMINDDH76 system was excellent for all CBC and white cell differential parameters (d% was comparable and even lower than the state of the art ones) (8, 20).

The carryover values in this analytical performance study were better compared with those recommended by the DYMINDDH76 manufacturer: the carryover for WBC, RBC, and HGB must be 0.5% and for PLT <1%. These data showed that there was not any transfer of material from one sample to the next sample, and the influence of contamination was close to zero. A lot of published studies recommend making a wash action between one sample and the next one to avoid a background count (9–11).

Linearity is an important measurement in the evaluation process. There should be a linear relation over a large concentration range at various dilutions for the parameter that is determined (1, 5, 11). In the present study, our attention was narrowed to low

thresholds of hematological parameters, where the clinicians have to make clinical decisions about the diagnosis and the appropriate treatment (whole blood transfusion, platelet transfusion) (13–15). The results showed excellent linear relationships between theoretical and obtained values with regression coefficients $r=0.999-1.000$ for all of the analyzed parameters.

When comparing two hematology devices, the variability is usually greater than in intra-patient variation, mainly because analyzers are independent analytical systems, often with different methods of determination, as well as with calibrators and quality control materials from various manufacturers (1, 15, 16).

Our data indicated a very good positive correlation between CBC and white cell differential, obtained by the DYMIND DH 76 system and results obtained by the SYSMEX XN 1000 system. Most correlation coefficients for reported parameters were greater than 0.95. Difference plot estimation of blood count and white cell differential count results from the DYMIND DH 76 system compared with results from the SYSMEX XN 1000 system shows no significant discrepancy between the methods. The slightly lower correlation coefficient for platelet count in samples with PLT less than $50 \times 10^9/L$ (0.910) was found. According to the published studies, the accuracy of automated platelet counting depends on the applied automated method based on impedance, optical, fluorescent optical principles or flow-cytometry (6, 7, 13). The SYSMEX XN 1000 system was able to use three of them, but for better comparison with DYMIND DH76, we performed the impedance method only.

As expected, cell types that occur in lower numbers in peripheral blood had a lower correlation coefficient compared to the other parameters included in the differential formula. The lower correlation coefficient for % basophils (0.855) was most likely due to

the use of a more sensitive fluorescent method on the SYSMEX XN 1000 system. An even lower correlation coefficient for basophils was recently reported by Kaplan and co-workers for a different line of cell counters (11, 17).

Correlation of the white cell differentials obtained on the DYMIND DH 76 hematology analyzer with results of the reference method, manual 200-cell counts by 2 researchers, was also very good. The performance was very good for neutrophils and lymphocytes ($r>0.97$), but it was clearly lower for monocytes ($r=0.836$) and for basophils ($r=0.625$). The decreased correlation coefficient for basophils is connected with the higher precision of the blood cell analyzers for cell types that occur in lower numbers in circulating blood because of the greater numbers of cells counted. In contrast, the decreased correlation coefficient for monocytes is not related to lower cell count in peripheral blood. In general, monocytes are difficult to identify due to their large morphological variations (2, 18, 19).

Conclusion

Performance evaluation and comparative analyses of the automated DYMIND DH76 hematology system showed excellent analytical characteristics and laboratory results comparable to those obtained by the established laboratory practice analyzers. The DYMIND DH76 is useful for all laboratories, mainly in those with a large number of normal and abnormal samples. Its user-friendly workstation has complete patient data management, which is in compliance with good laboratory practices.

Conflict of interest statement

All the authors declare that they have no conflict of interest in this work.

References

1. Briggs C, Culp N, Davis B, d'Onofrio G, Zini G, Machin SJ. ICSH guidelines for the evaluation of blood cell analyzers including those used for differential leucocyte and reticulocyte counting. *Int J Lab Hematol* 2014; 36: 613–27.
2. Verbrugge SE, Huisman A. Verification and standardization of blood cell counters for routine clinical laboratory tests. *Clin Lab Med* 2015; 35: 183–96.
3. Clinical and Laboratory Standards Institute (CLSI). Validation, verification, and quality assurance of automated hematology analyzers. Approved standard. 2nd ed. Wayne, PA: CLSI Document H26-A2: 2010.
4. Vis JY, Huisman A. Verification and quality control of routine hematology analyzers. *Int J Lab Hematol* 2016; 38: Suppl 1: 100–9.
5. Koepke AA, National Committee for Clinical Laboratory Standards. Reference leukocyte (WBC) differential count (proportional) and evaluation of instrumental methods: Approved Standard. H20-A2. Wayne, PA: CLSI; 2007.
6. Marionneaux SM, Maslak PG, Francisco N, Chan Y, Hanenberg J, Lynch J, et al. Potential clinical impact of inaccurate automated platelet counts in the setting of severe thrombocytopenia. *Blood* 2012; 120: 3428.
7. Briggs C, Kunka S, Machin SJ. The most accurate

- platelet count on the Sysmex XE 2100. Optical or impedance? *Clin Lab Haematol* 2004; 26: 157–8.
8. Buttarello M, Plebani M. Automated blood cell counts: state of the art. *Am J Clin Pathol* 2008; 130: 104–16.
 9. Nakul-Aquaronne D, Sudaka-Sammarcelli I, Ferrero-Vacher C, Starck B, Bayle J. Evaluation of the Sysmex XE-2100 hematology analyzer in hospital use. *J Clin Lab Anal* 2003; 17: 113–23.
 10. Maciel TS, Comar SR, Beltrame MP. Performance evaluation of the Sysmex® XE-2100D automated hematology analyzer. *J Bras Patol Med Lab* 2014; 50: 26–35.
 11. Kaplan SS, Johnson K, Wolfe N, Brown W, Keeney M, Gray-Statchuk L, Yee IC, et al. Performance characteristics of the Coulter LH 500 hematology analyzer. *Lab Hematol* 2004; 10: 76–87.
 12. McFarlane A, Aslan B, Raby A, Bourner G, Padmore R. Critical values in hematology. *Int J Lab Hematol* 2015; 37: 36–43.
 13. Segal HC, Briggs C, Kunka S, Casbard A, Harrison P, Machin SJ, et al. Accuracy of platelet counting haematology analysers in severe thrombocytopenia and potential impact on platelet transfusion. *Br J Haematol* 2005; 128: 520–5.
 14. Tanaka Y, Tanaka Y, Gondo K, Maruki Y, Kondo T, Asai S, et al. Performance evaluation of platelet counting by novel fluorescent dye staining in the XN-series automated hematology analyzers. *J Clin Lab Anal* 2014; 28: 341–8.
 15. Bruegel M, Nagel D, Funk M, Fuhrmann P, Zander J, Teupser D. Comparison of five automated hematology analyzers in a university hospital setting: Abbott Cell-Dyn Sapphire, Beckman Coulter DxH 800, Siemens Advia 2120i, Sysmex XE-5000, and Sysmex XN-2000. *Clin Chem Lab Med* 2015; 53: 1057–71.
 16. Becker PH, Fenneteau O, Da Costa. Performance evaluation of the Sysmex XN-1000 hematology analyzer in assessment of the white blood cell count differential in pediatric specimens. *Int J Lab Hematol* 2016; 38: 54–63.
 17. Meintker L, Ringwald J, Rauh M, Krause SW. Comparison of automated differential blood cell counts from Abbott Sapphire, Siemens Advia 120, Beckman Coulter DxH 800, and Sysmex XE-2100 in normal and pathologic samples. *Am J Clin Pathol* 2013; 139: 641–50.
 18. Okada T. Development and problem of automated hematology analyzer. *Sysmex J Int* 1999; 9: 52–7.
 19. Seo JY, Lee ST, Kim SH. Performance evaluation of the new hematology analyzer Sysmex XN-series. *Int J Lab Hematol* 2015; 37: 155–64.
 20. Vis JY, Huisman A. Verification and quality control of routine hematology analyzers. *Int J Lab Hematol* 2016; 38 Suppl 1: 100–9.

Received: October 17, 2020

Accepted: February 03, 2021